Prod		
Stent Material:	Cobalt Chromium L605	
Polymer:	Biodegradable and Biocompatible polymer (PLA and PLGA)	
Drug:	Sirolimus	
Drug Dosage:	1.33 μg/mm²	
Strut Thickness:	65 μm (0.065 mm or 0.0026")	
Nominal Pressure:	9 atm	
Rated Burst Pressure:	16 atm, 14 atm with 3.50mm diameter with length >35 mm and 4.00mm diameter with length >30mm and 4.50 mm diameter	
Mean Recoil:	Average 3.2%	•
Mean Foreshortening:	± 1.3 %	

ORDERING INFORMATION

Ø (mm)/ Length (mm)	8mm	13mm	16mm	19mm	24mm	29mm	32mm	37mm	40mm
2.00mm	DC2.1 2.0008	DC2.1 2.0013	DC2.1 2.0016	DC2.1 2.0019	DC2.1 2.0024	DC2.1 2.0029	DC2.1 2.0032	DC2.1 2.0037	DC2.1 2.0040
2.25mm	DC2.1 2.2508	DC2.1 2.2513	DC2.1 2.2516	DC2.1 2.2519	DC2.1 2.2524	DC2.1 2.2529	DC2.1 2.2532	DC2.1 2.2537	DC2.1 2.2540
2.50mm	DC2.1 2.5008	DC2.1 2.5013	DC2.1 2.5016	DC2.1 2.5019	DC2.1 2.5024	DC2.1 2.5029	DC2.1 2.5032	DC2.1 2.5037	DC2.1 2.5040
2.75mm	DC2.1 2.7508	DC2.1 2.7513	DC2.1 2.7516	DC2.1 2.7519	DC2.1 2.7524	DC2.1 2.7529	DC2.1 2.7532	DC2.1 2.7537	DC2.1 2.7540
3.00mm	DC2.1 3.0008	DC2.1 3.0013	DC2.1 3.0016	DC2.1 3.0019	DC2.1 3.0024	DC2.1 3.0029	DC2.1 3.0032	DC2.1 3.0037	DC2.1 3.0040
3.50mm	DC2.1 3.5008	DC2.1 3.5013	DC2.1 3.5016	DC2.1 3.5019	DC2.1 3.5024	DC2.1 3.5029	DC2.1 3.5032	DC2.1 3.5037	DC2.1 3.5040
4.00mm	DC2.1 4.0008	DC2.1 4.0013	DC2.1 4.0016	DC2.1 4.0019	DC2.1 4.0024	DC2.1 4.0029	DC2.1 4.0032	DC2.1 4.0037	DC2.1 4.0040
4.50mm	DC2.1 4.5008	DC2.1 4.5013	DC2.1 4.5016	DC2.1 4.5019	DC2.1 4.5024	DC2.1 4.5029	DC2.1 4.5032	DC2.1 4.5037	DC2.1 4.5040



Purple Microport Cardiovascular Private Limited 615/616 Avior, L.B.S. Road, Mulund (W), Mumbai-400080

Tel: +(91) 22 25916811/12 Email: info@purplemicro

Email: info@purplemicroport.com Website : www.purplemicroport.com

All cited trademarks are the property of their respective owners in respective countries. These devices shall only be used for curing diseases on prescription by a licensed physician. Indications contraindications, warnings, potential adverse events and directions for use can be found in instructions for use and product labeling supplied with each device. These devices are prohibited to promote, distribute or sell in any countries or regions where the registration certificates or any other licenses are not obtained. This brochure is not intended for use in aforesaid countries or regions. This brochure is only delivered to specific person for his or her own reference, is not to be distributed widely as an advertisement. Any information contained in this brochure may be changed without prior notice.

BPSS/BRO/20210901_05



2460



Innovative Stent Design with Flexibility and Strength

Biodegradable Polymer Based Sirolimus Eluting Coronary Stent System

